

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AMH-S-1-C-PN-M1*

S.N.:

3945

Date:

SEPT 23 87



a) Output signal amplitude:

0 TO ± 10 VOLTS

b) Pulse width:

≤ 130 PSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

120 / 240 V

50 - 60 Hz

40 dB ATTEN
10 VOLTS/DIV
1.0 MHz

M1: FRONT PANEL
MOUNTED PN