


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-5-1-P-e

S.N.: 3973

Date: NOV 11 1987



40 db ATTEN  
10 VOLT/DIV  
1.0 MHz

- a) Output signal amplitude:  
0 TO +10 VOLTS
- b) Pulse width:  
≤ 130 PSEZ
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

120/240 V  
50-60 Hz

*[Signature]*