


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1-75-CA-EN-ED-P*

S.N.: *3978*

Date: *NOV 9 1987*



*40 db ATTEN
10 VOLTS/DIV
1.0 ns/Hz*

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240V
50-60 Hz*