

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-1-PW-PS-P*
S.N.: *3979*
Date: *OCT 30 87*



R_L = 50 Ω
100 KHz.

- a) Output signal amplitude:
0 TO +200 V
- b) Pulse width:
0.1 TO 5.0 μ SEC
- c) Rise time:
 \leq 10 NSEC
- d) Fall time:
 \leq 10 NSEC
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz.