


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-1-PW-C-2*

S.N.: *3986*

Date: *OCT 27 87*



*R<sub>L</sub> = 50Ω*  
*1 kHz*

- a) Output signal amplitude:  
*0 TO 7200 VOLTS*
- b) Pulse width:  
*0.1 TO 5.0 USEC.*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC.*
- e) PRF:  
*0 TO 1 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240V*  
*50-60 Hz*  
*[Signature]*