


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-TN-1
S.N.: 4003
Date: SEPT 16 87



60 dB ATTEN
100 VOLTS/DIV
1 KHz.

- a) Output signal amplitude:
-230 VOLTS
of +30 VOLT OFFSET
- b) Pulse width:
5 TO 100 NSER
- c) Rise time:
2 NSER
- d) Fall time:
≤ 3 NSER
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz
- h) PROPAGATION DELAY:
≤ 100 NSER.