



PULSE GENERATOR  
PERFORMANCE CHECK

Model: APR-3 PW-PS-T  
S.N.: 4021  
Date: SEPT 25 87



Ⓐ  $R_L = 50 \Omega$   
 $500 \text{ Hz}$



Ⓑ  $R_L = 50 \Omega$   
 $50 \text{ Hz}$

- a) Output signal amplitude:  
0 TO +200 V.
- b) Pulse width:  
0.1 TO 100 nSEC
- c) Rise time:  
 $\leq 10 \text{ nSEC}$
- d) Fall time:  
 $\leq 10 \text{ nSEC}$
- e) PRF:  
MIN PW: 0 TO 1 KHz  
MAX PW: 0 TO 50 Hz.
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 Hz.