


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-M-1-P*

S.N.: *4025*

Date: *NOV 23 1967*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*40 db ATTEN
10 VOLTS DIV
1.0 MHz*

*+15 ± 10 VOLTS
250 mA (MAX)*