


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV4-MPT-P*

S.N.: *4033*

Date: *DEC 2 1987*



- a) Output signal amplitude:  
*≥ 30 VOLTS*
- b) Pulse width:  
*≤ 1.0 NSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*7.15 ± 1.0 VOLTS*  
*150 mA (MAX)*

*50 db ATTEN*

*32 VOLTS/DIV*

*500 KHz.*

*AB*