


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-P .

S.N.: 4035

Date: DEC 2, 1987



50 db ATTEN
∴ 32 VOLTS/DIV
500 KHZ

- a) Output signal amplitude:
 ≥ 30 VOLTS
- b) Pulse width:
 ≤ 1.0 NSEC
- c) Rise time:
 NA
- d) Fall time:
 NA
- e) PRF:
 0 TO 1 MHZ
- f) Jitter, stability:
 OK
- g) Prime power:
 $+15 \pm 1.0$ VOLT
 150 mA (MAX)