

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVL-2-C-P

S.N.:

4044

Date:

SEPT 22 87



70 db atten
2.320 volts/div
2.0 kHz

a) Output signal amplitude:

0 TO +350 V.

b) Pulse width:

5 TO 100 nsec.

c) Rise time:

1.2 nsec

d) Fall time:

1.3 nsec.

e) PRF:

0 TO 5 kHz

f) Jitter, stability:

OK

g) Prime power:

120 / 240 V

50-60 Hz.

[Signature]