


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AP-1-M-CA-N*

S.N.: *4056*

Date: *DEC 17 87*



40 db ATTEN
10 VOLTS DIV
1.0 MHz

- a) Output signal amplitude:
0 TO 10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 35 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT
300 mA (MAX).