


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-3-C-7*

S.N.: *4066*

Date: *JAN 17 '88*



*70 dB ATTEN
≈ 320 VOLTS/DIV
20 KHz*

- a) Output signal amplitude:
0 TO +500 VOLTS
 - b) Pulse width:
65 NS (HALF-PEAK)
 - c) Rise time:
NA
 - d) Fall time:
NA
 - e) PRF:
0 TO 20 KHz
 - f) Jitter, stability:
OK
 - g) Prime power:
*120/240V
50-60 Hz*
- [Signature]*