


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMH-5-1-C-N*

S.N.: *4070*

Date: *NOV 5 1987*



Ⓐ 40 dB ATTN
10 VOLTS/DIV
1.0 MHz



Ⓑ AS Ⓐ BUT
200 pSEC/DIV

- a) Output signal amplitude:
0 TO -10 VOLTS?
- b) Pulse width:
≤ 130 pSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz