


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PAR-A-1-PW-C-P*

S.N.: *4071*

Date: *APR 8 '88*



50 VOLTS/DIV

R_L = 50 Ω

a) Output signal amplitude:

0 TO +200 V

b) Pulse width:

50 nSEC TO 1.0 μ SEC

c) Rise time:

≤ 10 nSEC

d) Fall time:

≤ 10 nSEC

e) PRF:

MIN PW:

f) Jitter, stability:

OK

g) Prime power:

120 / 240V

50-60 Hz.

[Signature]