


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PP-10-1-C-M-PN*

S.N.: *4078* *1*

Date: *DEC 1 87*



40 dB ATTEN

1.10 VOLTS/DIV

1.00 nS

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0.2 TO 4.0 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz