


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PVO-5B-C-PN*

S.N.: *4083*

Date: *JAN 14 1988*




- a) Output signal amplitude:  
*0 TO ± 48 AMPS*
- b) Pulse width:  
*25 TO 500 NSEC*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:

Power

*$R_L = 0.7 \Omega$   
 $PRF = 1.0 \text{ KHz}$*

- g) Prime power: *OK*  
*120/240V*  
*50-60 Hz*



Notes  *$R_L = 0.7 \Omega$   
 $PRF = 1.0 \text{ KHz}$*

*[Signature]*