


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVR-3-C-P-CA-PW*

S.N.: *4093*

Date: *JAN 26 88*



① $R_L = 50\Omega$
 $PRF \approx 1.0 \text{ KHz}$



② $R_L = 50\Omega$
 $PRF = 50 \text{ Hz}$

- a) Output signal amplitude:
0 TO +200.40V
- b) Pulse width:
0.1 TO 100 μ SEC
- c) Rise time:
 $\leq 10 \text{ NSEC}$
- d) Fall time:
 $\leq 10 \text{ NSEC}$
- e) PRF:
TRAIN PW TO 1 μ SEC: 0 TO 1KHz
MAX PW: 0 TO 50 Hz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]