


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV14-MPT-P*

S.N.: *4094*

Date: *DEC 3 1987*



- a) Output signal amplitude:
> 30 VOLTS.
- b) Pulse width:
≤ 1 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*50 db ATTEN
32 VOLTS/DIV
500 KHz.*

*+15 ± 1.0 VOLTS
150 mA (MAX)*