


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-P.

S.N.: 4095

Date: DEC 3 1987



50 db ATTEN
∴ 32 VOLTS/DIV
500 KHZ

- a) Output signal amplitude:
 $\geq 30 \text{ VOLTS}$
- b) Pulse width:
 $\leq 1 \text{ NSEC}$
- c) Rise time:
 N/A
- d) Fall time:
 N/A
- e) PRF:
 0 TO 1MHZ
- f) Jitter, stability:
 OK
- g) Prime power:
 +15 \pm 1.0 VOLTS
 150 mA (MAX)