


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-P

S.N.: 4097

Date: DEC 3 1987



50 db ATTEN
∴ 32 VOLTS/DIV
500 KHZ

- a) Output signal amplitude:
 ≥ 30 VOLTS
- b) Pulse width:
 ≤ 1 NSEC
- c) Rise time:
 NA
- d) Fall time:
 NA
- e) PRF:
 0 TO 1MHZ
- f) Jitter, stability:
 OK
- g) Prime power:
 $\pm 15 \pm 1.0$ VOLTS
 150 mA (MAX)