


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVC-TN-1*

S.N.: *4100*

Date: *NOV 10 1987*



- a) Output signal amplitude:
-230 VOLTS
- b) Pulse width:
WITH +40V DC OFFSET
5 TO 100 NSEC
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
53 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN

100 VOLTS/DIV

2.0 KHZ

120/240V

50-60 Hz

[Signature]