


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVL-TN 1*  
S.N.: *4101*  
Date: *NOV 10 1987*



- a) Output signal amplitude:  
*-230 VOLTS*  
*WITH +40 V DC OFFSET*
- b) Pulse width:  
*5 TO 100 NS OR*
- c) Rise time:  
*5 2 NS OR*
- d) Fall time:  
*4 3 NS OR*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*60 dB ATTEN*  
*100 VOLTS/DIV*  
*2.0 KHZ*

*120/240V*  
*50-60 HZ*  
*[Signature]*