


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVO-1-CP

S.N.: 4123

Date: JAN 21 88



- a) Output signal amplitude: +25 TO +100 AMPS
- b) Pulse width: ≤ 25 NSEC
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1 KHZ
- f) Jitter, stability: OK

≈ 60 AMPS/DN
1.0 KHZ
120/240 V
50-60 HZ

[Signature]