


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVM-2-N

S.N.: 4144

Date: Dec. 24, 1987



40 db atten.
10V / div
5.0 MHz

- a) Output signal amplitude:
0 to -15V
- b) Pulse width:
0.2 to 2.0 nsec
- c) Rise time:
 ≤ 100 psec
- d) Fall time:
 ≤ 135 psec
- e) PRF:
0 to 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 \pm 1.0 Volt
500 mA (max.)