


PULSE GENERATOR
PERFORMANCE CHECK


Model: PMM-3 -C -P

S.N.: 4150

Date: JAN 26 88



Ⓐ 40 dB ATTEN
10 VOLTS/DIV
≈ 5.0 MHz



Ⓑ AS Ⓐ BUT
≈ 25 MHz

- a) Output signal amplitude:
0 TO +15 V.
- b) Pulse width:
2 TO 15 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC.
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

[Handwritten signature]