


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVRL-ITT 7F-05-ED

S.N.: 4161

Date: FEB 10 1988



A out
60 db ATTEN
100 VOLTS/DIV



B out
 $R_L = 10K.$

- a) Output signal amplitude:
 - A) -200 V TO 50
 - B) -200 V TO > 1.0 K
- b) Pulse width:
 - A) 5 TO 10 NSER
 - B) 10 TO 150 NSER
- c) Rise time:
 - A) ≤ 3 NSER
 - B) ≤ 5 NSER
- d) Fall time:
 - SEE RISE TIME
- e) PRF:
 - 0 TO 1 KHz.
- f) Jitter, stability:
 - OK
- g) Prime power:
 - 120 / 240 V
 - 50-60 Hz
- h) OFFSET:
 - A) 0 TO +50 V
 - B) 0 TO +50 V.

