


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-2-C-N-05-M*  
S.N.: *4166*  
Date: *JAN 15 88*



*≈ 6 AMPS/DIV*  
*PRF = 10 KHz*

- a) Output signal amplitude:  
*0 TO -10 AMPS*
- b) Pulse width:  
*2 TO 15 NSEC*
- c) Rise time:  
*≤ 1 NSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*

*[Signature]*