


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV14-HV1-P*

S.N.: *4173*

Date: *MAR 10 1988*



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
≤ 1.0 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS.

60 dB ATTEN
≈ 100 VOLTS/DIV
50 KHz (PRF)

250 mA (max)
[Signature]