


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMH-MPT-P*

S.N.: *4223*

Date: *MAR 9 80*



- a) Output signal amplitude:  
*> +30 VOLTS*
- b) Pulse width:  
*≤ 1 NSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 10 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 db ATTEN*  
*32 VOLTS/DIV*  
*500 KHz*

*+15 ± 1.0 VOLTS*  
*200 mA MAX*  
*✓*