


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-MPT-P*

S.N.: *4226*

Date: *MAR 9 1988*



- a) Output signal amplitude:  
*71 ± 30 V<sub>0.8</sub>*
- b) Pulse width:  
*≤ 10 nsec.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*15 ± 1.0 V<sub>0.8</sub>*  
*200 mA (max)*

*50 db ATTEN*  
*32 VOLT/DIV*  
*500 KHz*