


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2A-CW-P

S.N.: 4231

Date: MAR 25 1988



- d) Output signal amplitude:  
0 TO +170 VOLTS
- e) Pulse width:  
3 TO 400 NSER \*
- f) Rise time:  
≤ 2 NSER
- g) Fall time:  
≤ 2 NSER
- h) PRF:  
0 TO 5 KHZ
- i) Jitter, stability:  
OK
- j) Prime power:

60 db ATTEN  
100 VOLTS/DIV  
5.0 KHZ  
 $R_L = 50 \Omega$

120/240 V  
30-60 HZ

\* NEEDS ADDITIONAL  
EXTERNAL CABLE  
FOR 100 TO 400 NSG  
PULSE.