



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-14V3-C-OT-PN*
S.N.: *4236*
Date: *APRIL 25 1988*



Ⓐ RISE TIME
60 db ATTEN
= 100 VOLTS/DIV



Ⓑ AS Ⓐ BUT
5 NSEC/DIV

- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
6 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/1240 V
50-60 Hz

[Signature]