


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2-N*

S.N.: *4250*

Date: *APRIL 7 1988*



- a) Output signal amplitude:  
*0 TO 350 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*70 db ATTEN*  
*≈ 320 VOLTS/DIV*  
*2 KHz*

*+15 ± 1.0 VOLT*  
*500 mA (MAX)*  
*[Signature]*