


PULSE GENERATOR
PERFORMANCE CHECK


Model: PCR-1-C-PN-CA-EW-SNL1

S.N.: 4259

Date: MAY 31 88



Ⓐ $P_{out} R_L = 33 \Omega$
 1 kHz



Ⓑ $N_{out} R_L = 33 \Omega$
 1 kHz

- a) Output signal amplitude:
0 TO $\pm 250 \text{ mV}$
- b) Pulse width: TO 40 ns
50 TO 500 ns
- c) Rise time:
 $\leq 10 \text{ ns}$
- d) Fall time:
 $\leq 10 \text{ ns}$
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
 $120/240 \text{ V}$
 $50-60 \text{ Hz}$