


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-1TT2-05*

S.N.: *4276*

Date: *MAR 4 1988*



- a) Output signal amplitude:
-200 V TO $R_L > 1k$
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
 ≤ 2 nSEC
- d) Fall time:
 ≤ 3 nSEC
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*≈ 100 VOLTS/DIV
500 KHz PRF.*

*120/240 V
50-60 Hz
h) OFFSET:
0 TO ± 50 VOLTS*