


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-2-PS-P*

S.N.: *4289*

Date: *MAY 4 1988*



- a) Output signal amplitude:  
*0 TO 710 AMPS*
- b) Pulse width:  
*3 TO 15 NSEC*
- c) Rise time:  
*≤ 1 NSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*≈ 7 AMPS/DIV*  
*20 KHz.*