


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-V-HV3-C-N*

S.N.: *4291*

Date: *MAR 30 88*



*(A) 60 db ATTEN
100 VOLTS/DIV*



*(B) 1nS (A) BUT
1 nSEC/DIV
(RISE TIME)*

- a) Output signal amplitude:
0 TO 100 VOLTS
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
≤ 7 nSEC
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50 - 60 Hz*

[Signature]