


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1-P*
S.N.: *4293*
Date: *MAY 10 1988*



40 db ATTEN
10 VOLTS/DIV

500 KHZ PRF. +15 ± 1.0 VOLTS
300 mA (MAX).

- a) Output signal amplitude:
0 TO 110 VOLTS.
- b) Pulse width:
0.2 TO 4.0 NSER
- c) Rise time:
≤ 100 PSEK
- d) Fall time:
≤ 135 PSEK
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

[Signature]