


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1123-P*

S.N.: *4304*

Date: *MAY 10 1988*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 TO 2.0 nSEC
- c) Rise time:
≤ 150 pSEC
- d) Fall time:
≤ 250 pSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS
300 mA (MAX)

(P)

