


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MM-1-P*
S.N.: *4306*
Date: *MAY 13 1988*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*40 db return
5 VOLTS/DIV
7.0 MHz.*

*+24 ± 1.0 VOLTS
500 mA (MAX)*
[Signature]