


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-C-P-1A*

S.N.: *4313*

Date: *MAY 12 1948*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0.2 TO 2.0 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz

40 db ATTEN
10 VOLTS/DIV
≈ 12 MHz. POST-2

[Signature]