


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-MU-14V3-C-P-M.*

S.N.: *4335*

Date: *APRIL 27 '88*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 TO 2.0 nSEC.
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 250 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

50 db ATTEN
≈ 32 VOLTS/DIV
1.0 MHz PRF

120/240 VOLTS
50-60 Hz
[Signature]