

PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-V-1123-C-M-N*

S.N.: *4336*

Date: *MAY 30 1988*



Ⓐ 60 db ATTEN
2 100 VOLTS/DIV
5.0 kHz



Ⓑ AS Ⓐ BUT 1 nSEC/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO -100 VOLTS
- b) Pulse width:
4 TO 100 nSEC
- c) Rise time:
≤ 200 pSEC
- d) Fall time:
≤ 4 nSEC
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240
50 - 60 Hz

[Signature]