


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-U-1A23-P*

S.N.: *4337*

Date: *MAY 4 1988*



Ⓐ *60 dB ATTEN*
100 VOLTS/DIV
5 KHz



Ⓑ *AS Ⓐ BUT*
1 nSEC/DIV
(RISETIME)

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 200 pSEC
- d) Fall time:
≤ 4 nSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLTS
300 mA (MAX.)