


PULSE GENERATOR
PERFORMANCE CHECK


Model: *ACMM-3-C-P-TR*

S.N.: *4343*


Date: *JUNE 3 1988*



Ⓐ OUT PORT
40 dB ATTEN
10 VOLTS/DIV
12 MHz



Ⓑ AS Ⓐ BUT TR OUT PORT
WIDER RISE TIME



Ⓒ AS Ⓑ BUT *MUR*
RISE TIME

- a) Output signal amplitude:
0 TO 770 VOLTS
- b) Pulse width:
0.5 TO 5.0 NSEC
- c) Rise time:
0.3 TO 2.0 NSEC
- d) Fall time:
0.6 TO 2.0 NSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50-60 Hz