


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-C-N-M-EN

S.N.: 4350

Date: MAY 26 1988



- b) Output signal amplitude:  
0 to 350 VOLTS
- b) Pulse width:  
5 to 100 NSEC
- c) Rise time:  
≤ 2 NSEC
- d) Fall time:  
≤ 3 NSEC
- e) PRF:  
0 to 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

~ 70 db ATTEN

320 VOLTS/DIV

100 KHz

120 / 240 V

50 - 60 Hz