


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-M1-P-40*

S.N.: *4359*

Date: *JUNE 8 1988*



- a) Output signal amplitude:
≥ +40 VOLTS
- b) Pulse width:
≤ 1.0 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN

10 VOLTS/DIV

700 KHz PRF

+135 ± 5.0 VOLTS

5 mA (MAX).