


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-MI-P-40.

S.N.: 4367

Date: JUNE 8, 1988



- a) Output signal amplitude:  
 $\geq +40$  VOLTS
- b) Pulse width:  
 $\leq 1.0$  NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
 $+135 \pm 5.0$  VOLTS  
5 mA (MAX)

.40 db ATTEN

10 VOLTS/DIV

700 KHz PRF