



PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-A-1-PW-C-EXT-EN-TN*
 S.N.: *4380*
 Date: *JUL 15 1988*



- a) Output signal amplitude: *0 TO 1200 VOLTS*
- b) Pulse width: *50 NSER TO*
- c) Rise time: *1.0 USER*
6 10 NSER
- d) Fall time: *6 10 USER*
- e) PRF: *0 TO 100 KHz*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz

Ⓐ *P out*
R_i = 50 Ω
PRF = 5 KHz



Ⓑ *N out*
R_i = 50 Ω
PRF = 50 KHz

[Signature]